



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN™ Protein C

Item Code: 120003

Lot No.: FE33611C / FE3362

Expiration date: 2027-01-06

Components	Quantity	Lot	Expiration Date
R1: Reaction Buffer	4 vials	FE3361en1C / FE3362	2027-01-06
R2: Latex reagent	4 vials	FE3361eo1C / FE3362	2027-02-04

Analytical Data	Specification	Result
Calibration curve : reactivity, slope and accuracy	OD/min (3/2) \geq 0.30 Delta OD/min (1/1-1/2) \geq 0.1 Controls measured within acceptance ranges	0.41 0.13 PASS
Hook effect	OD/min (300 et 600 %) \geq OD/min (point 1/1 of calibration curve)	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval: **Quality Control Approval**

21 OCT. 2025

S. LECOURT

Quality Control Manager: