



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN a2 ANTIPLASMIN (LRT)
Product Code : 220502
Lot No. : FE03781H
Expiration date : 2027-01-31

Components	Quantity	Lot	Expiration Date
R1 : Plasmin	3 vials	FE0378171H	2027-01-31
R2 : SPm-41 Substrate	3 vials	FE0378181H	2027-01-31

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05 A405/min(100%) = 0.25 ± 0.10 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.085 0.239 PASS PASS
Detection threshold (LOQ) on CS	$\leq 10\%$	< 3%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

03 MARS 2025
S. LECOURT

Quality Control Manager :