



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL PLASMA
Product Code : 223101
Lot No. : FD2373
Expiration date : 2027-01-09

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD1434BXML	2027-01-09
C2	6 vials	FD1423BYLK	2027-01-09

Analytical Data	Specification	Result
Concentration	C1 : 0.10 -0.30 IU/ml C2 : 0.40 -0.60 IU/ml	0.18 IU/mL 0.50 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**
18 NOV. 2024
S. LECOURT

Quality Control Manager :

