



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS	
Product Description :	BIOPHEN APIXABAN CONTROL
Product Code :	225301
Lot No. :	FD18691F
Expiration date :	2027-01-23

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD18691F	2027-01-23
C2	6 vials	FD18691F	2027-01-23

Analytical Data	Specification	Result
Concentration	C1 : 140 -260 ng/ml C2 : 340 -460 ng/ml	190 ng/mL 416 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :



Quality Control Approval

17 OCT. 2024

S. LECOURT