



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** COLLAGEN  
**Product Code :** AG005K  
**Lot No. :** FD1551LY / FD1548  
**Expiration date :** 2027-01-25

Components	Quantity	Lot	Expiration Date
COLLAGEN REAGENT	3 vials	FD1551LY / FD1548	2027-01-25
COLLAGEN DILUENT	3 vials	FD1551LY / FD1548	2027-01-29

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 2 µg/ml on CS or CN Series	≥ 70%	85%
Reactivity (% AgregMax) at 2 µg/ml on Agregometer	≥ 70%	84%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

23 SEP. 2024

S. LECOURT