



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN FACTOR X

**Product Code :** 221705

**Lot No. :** FD17571C

**Expiration date :** 2027-02-26

Components	Quantity	Lot	Expiration Date
R1 : Substrate	4 vials	FD17571C	2027-03-09
R2 : RVV	4 vials	FD17571C	2027-03-15
R3: Buffer	4 vials	FD17571C	2027-02-26

Analytical Data	Specification	Result
Calibration curve : reactivity, linearity, slope and accuracy	$A_{405} (1:5) \geq 0.75$ $\Delta A_{405} (1:5 ; 1:10) \geq 0.35$ $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.07 0.52 1.00 PASS
Detection Threshold	$\leq 5\%$	< 1 %

### Calibration Curve for information

STAR

Conc.	0	50	100	200
OD	0.00	0.27	0.55	1.07

	Acceptance Range	MV
NC	[ 69 - 89 ]	82
AC	[ 27 - 36 ]	33

MV = Measured value

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval** Quality Control Manager :

15 OCT. 2024

S. LECOURT