



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR
Product Code : 226501
Lot No. : FD16871F
Expiration date : 2027-02-20

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD1687aa1F	2027-02-20
CAL 2	4 vials	FD1687ab1F	2027-02-20
CAL 3	4 vials	FD1687ac1F	2027-02-20

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 40 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/ml 244 ng/ml 474 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

24 SEP. 2024

S. LECOURT