



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : CEPHEN 2.5 LS
Product Code : CK522K
Lot No. : FE03812T / FE043835
Expiration date : 2027-02-16

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5 LS	6 vials	FE0381KD2T / FE043835	2027-02-16

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.04 on STAR 1.05 on CS 1.04 on CN

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

28 MARS 2025

S. LECOURT