



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR
Product Code : 222701
Lot No. : FD19741F
Expiration date : 2027-03-30

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FD1974BA1F	2027-03-30
CAL 2	4 vials	FD1974BB1F	2027-03-30
CAL 3	4 vials	FD1974BC1F	2027-03-30

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -300 ng/ml CAL 3 : 400 -500 ng/ml	0 ng/ml 253 ng/ml 460 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

22 OCT. 2024
S. LECOURT

