



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN FXIa CONTROL SET
Product Code : 224801
Lot No. : FD19891A
Expiration date : 2027-03-28

Components	Quantity	Lot	Expiration Date
C1	3 vials	FD19891A	2027-03-28
C2	3 vials	FD19891A	2027-03-28

Analytical Data	Specification	Result
Concentration	C1 : about 10 mIU/ml C2 : about 30 mIU/ml	13 mIU/mL 34 mIU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

29 OCT. 2024

S. LECOURT