



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ Rivaroxaban Control Low
Product Code : 225101
Lot No. : FD18791H
Expiration date : 2027-03-23

Components	Quantity	Lot	Expiration Date
CI	6 vials	FD1879DM1H	2027-03-23
CII	6 vials	FD1879DN1H	2027-03-23

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -75 ng/ml	27 ng/ml 65 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :
21 OCT. 2024
S. LECOURT