

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : IMIDAZOLE BUFFER
Product Code : AR021K/L
Lot No. : FD1796MS
Expiration date : 2027-03-05

Components	Quantity	Lot	Expiration Date
IMIDAZOLE BUFFER	3 vials *25mL (AR021K) 12 vials *25mL (AR021L)	FD1796ARMS	2027-03-05

Analytical Data	Specification	Result
Hemoclot PC clotting assay	Delta CT (100%) with control sample \leq 5%	<1%
FVIII:C clotting assay	Delta CT (100%) with control sample \leq 5%	<1%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

30 SEP. 2024
S. LECOURT

Quality Control Manager :

