



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW CII
Product Code : 224401
Lot No. : FD2463PV / FD2464PW
Expiration date : 2027-05-24

Components	Quantity	Lot	Expiration Date
CII	12 vials	FD2463BRPV / FD2464PW	2027-05-24

Analytical Data	Specification	Result
Concentration	CII : 0.35 -0.65 IU/ml	0.50 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval Quality Control Manager :

02 JAN. 2025

S. LECOURT