



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

CERTIFICATE OF ANALYSIS	
Product Description :	Calcium Chloride 0.025M
Product Code :	AR001B/L
Lot No. :	FD2332PD
Expiration date :	2027-05-07

Components	Quantity	Lot	Expiration Date
Calcium Chloride 0.025M	3 vials (AR001B) 12 vials (AR001L)	FD2332j2PD	2027-05-07

Analytical Data	Specification	Result
aPTT on citrated human plasma (CEPHEN)	ratio $\leq$ 1.20	1.03
FVIII:C clotting assay	Delta CT (100%) with control sample $\leq$ 5%	< 1%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

**Quality Control Approval**

**02 DEC. 2024**

**S. LECOURT**

Quality Control Manager :