



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN™ Free Protein S

Item Code : 120004

Lot No. : FE26501M / FE2656

Expiration date : 2027-06-17

Components	Quantity	Lot	Expiration Date
R1 : Reaction Buffer	4 vials	FE2650bu1M / FE2656	2027-06-17
R2 : Latex reagent	4 vials	FE2650bv1M / FE2656	2027-06-17

Analytical Data	Specification	Result
Calibration curve : reactivity, slope and accuracy	OD/min (point 3/2) ≥ 0.09 Delta OD/min (11/10 - 3/8) ≥ 0.04 Controls measured within acceptance ranges	0.15 0.08 PASS
Specificity	$\leq 10\%$ Free PS:Ag	<1%
Hook effect	OD/min (600%) \geq OD/min (point 3/2 of calibration curve)	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

15 JUL. 2025

S. LECOURT