



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : CEPHEN 2.5
Product Code : CK512K
Lot No. : FE3273RN / FE3278
Expiration date : 2027-06-27

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5	6 vials	FE3273KBRN / FE3278	2027-06-27

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.00 on STAR 0.99 on CS 0.99 on CN

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval
Date of Approval : 08 SEP. 2025
S. LECOURT
Quality Control Manager :