



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor IX Deficient Plasma
Product Code : DP050A/K
Lot No. : FD2565Q1 / FD2571Q2
Expiration date : 2027-06-29

Components	Quantity	Lot	Expiration Date
Factor IX Deficient Plasma	1 vial (DP050A) 6 vials (DP050K)	FD2565JWQ1 / FD2571Q2	2027-06-29

Analytical Data	Specification	Result
Calibration curve	CT (0%) $\geq$ 40 SEC CT (100%) $<$ 80 SEC	162sec 54sec
Concentration FIX	$< 1\%$	$< 1\%$
Concentration FV	$> 50\%$	85%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **30 JAN. 2025**

Quality Control Manager :

Sophie LECOURT
Deputy Director Quality Control