



HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN NORMAL CONTROL PLASMA

Product Code : 223201

Lot No. : FE00741E / FE02412L

Expiration date : 2027-07-19

Components	Quantity	Lot	Expiration Date
Normal Control	12 vials	FE0074DZ1E / FE02412L	2027-07-19

Analytical Data	Specification	Result
INR	<1.3	<1.30
aPTT LR Ratio	≤ 1.20	≤1.20
Fibrinogen (Activity)	> 2 g/L	2.68 g/L
Fibrinogen (Antigen)	> 2 g/L	2.80 g/L
Thrombin time Ratio	≤ 1.20	≤1.20
LA ratio	≤ 1.2	≤1.2
Alpha 2 Anti plasmin	≥ 80 %	89%
Antithrombin (Activity)	> 80 %	90 %
Antithrombin (Antigen)	> 80 %	89 %
Factor II	> 70 %	99 %
Factor V	> 70 %	103 %
Factor VII	> 70 %	103 %
Factor X	> 70 %	88 %
Factor VII+ X	> 70 %	95 %
Factor VIII	> 60 %	92 %
Factor IX	> 60 %	86 %
Factor XI	> 70 %	84 %
Factor XII	> 70 %	97 %
Factor XIII	> 60 %	90 %
Free Protein S (Antigen)	> 70 %	102 %
Plasminogen	> 75 %	99 %
Protein C Chromogenic	> 80 %	92 %
Protein C (Antigen)	> 70 %	82 %
Protein C Clotting	> 60 %	94 %
Protein S (Activity)	> 50 %	82 %
von Willebrand Factor (Antigen)	> 60 %	107 %
Act. PC-r (quantitative)	< 10 % FVL	2 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

30 AVR. 2025

S. LECOQ

