



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description BIOPHEN ABNORMAL CONTROL PLASMA

Product Code : 223301

Lot No. : FE007914 / FE02131A

Expiration date : 2027-07-11

Components	Quantity	Lot	Expiration Date
Abnormal Control	12 vials	FE0079EA14 / FE02131A	2027-07-11

Analytical Data	Specification	Result
INR	> 1,3	>1.3
aPTT LR Ratio	> 1.50	>1.50
Fibrinogen (Activity)	0.8-1.2 g/L	1.10 g/L
Fibrinogen (Antigen)	0.7-1.3 g/L	1.20 g/L
LA ratio	≤ 1.2	≤1.2
Alpha 2 Anti plasmin	20 - 50 %	34 %
Antithrombin (Activity)	20 - 50 %	35 %
Antithrombin (Antigen)	20 - 50 %	34 %
Factor II	20 - 50 %	38 %
Factor V	20 - 50 %	30 %
Factor VII	20 - 50 %	40 %
Factor X	20 - 50 %	35 %
Factor VII+ X	20 - 50 %	37 %
Factor VIII	20 - 50 %	31 %
Factor IX	20 - 50 %	33 %
Factor XI	20 - 50 %	29 %
Factor XII	20 - 50 %	37 %
Factor XIII	20 - 50 %	44 %
Free Protein S (Antigen)	20 - 50 %	32 %
Plasminogen	20 - 50 %	39 %
Protein C Chromogenic	20 - 50 %	34 %
Protein C (Antigen)	20 - 50 %	30 %
Protein C Clotting	20 - 50 %	38 %
Protein S (Activity)	20 - 50 %	38 %
von Willebrand Factor (Antigen)	20 - 50 %	42 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

30 AVR. 2025

S. LECOURT

Quality Control Manager :