



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** BIOPHEN™ BIVALIRUDIN CONTROL  
**Product Code :** 225701  
**Lot No. :** FE04281A / FE04271A  
**Expiration date :** 2027-07-30

| Components | Quantity | Lot                   | Expiration Date |
|------------|----------|-----------------------|-----------------|
| C1         | 3 vials  | FE0428de1A / FE04271A | 2027-07-30      |
| C2         | 3 vials  | FE0428df1A / FE04271A | 2027-07-30      |

| Analytical Data | Specification                                    | Result                             |
|-----------------|--------------------------------------------------|------------------------------------|
| Concentration   | C1 : 1.25 - 1.75 µg/ml<br>C2 : 2.50 - 3.30 µg/ml | C1 : 1.74 µg/mL<br>C2 : 3.16 µg/mL |

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

### Quality Control Approval

Date of Approval :

03 MARS 2025

S. LECOURT

Quality Control Manager :

