



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF: CBA

Product Code : RK038A

Lot No. : FD2354

Expiration date : 2027-07-26

Components	Quantity	Lot	Expiration Date
Collagen pre-coated plate	12x8 wells	FD2355	2028-05-01
vWF standard	3 vials	FE0112	2028-06-26
vWF Control I	1 vial	FE0110	2028-06-26
vWF Control II	1 vial	FE0111	2028-06-26
Anti-(h)- vWF HRP immunoconjugate	3 vials	FE0113	2028-06-29
Sample diluent	2x50 ml	FE0193	2028-07-03
Wash solution	1x50 ml	FD0331	2027-07-26
Conjugate diluent	1x25 ml	FD2352	2028-05-01
TMB	1x25 ml	241004D01	2028-10-31
Sulfuric Acid 0,45M	1x6 ml	FD0463	2027-08-09

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.00	2.05
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0
Detection threshold	$\leq 5\%$	2

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	2.410	1.717	1.102	0.5	0.273

	Acceptance Range	MV
CI	[62 - 83]	74
CII	[30 - 45]	34

MV = Measured value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 2025-02-06

Quality Control Manager : 