



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN HEPARIN CALIBRATOR
Product Code : 222001
Lot No. : FE05363J / FE05733P
Expiration date : 2027-08-21

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FE0536BJ3J / FE05733P	2027-08-21
CAL 2	4 vials	FE0536BK3J / FE05733P	2027-08-21
CAL 3	4 vials	FE0536BL3J / FE05733P	2027-08-21
CAL 4	4 vials	FE0536BM3J / FE05733P	2027-08-21
CAL 5	4 vials	FE0536BN3J / FE05733P	2027-08-21

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/ml CAL 2 : 0.30 -0.50 IU/ml CAL 3 : 0.70 -0.90 IU/ml CAL 4 : 1.10 -1.30 IU/ml CAL 5 : 1.45 -1.75 IU/ml	0 IU/mL 0.40 IU/mL 0.82 IU/mL 1.21 IU/mL 1.60 IU/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

18 MARS 2025

S. LECOURT

Quality Control Manager :

