



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ARIXTRA CALIBRATOR
Product Code : 222501
Lot No. : FE04001B / FE044113
Expiration date : 2027-08-03

Components	Quantity	Lot	Expiration Date
CAL 1	3 vials	FE0400CN1B / FE044113	2027-08-03
CAL 2	3 vials	FE0400CO1B / FE044113	2027-08-03
CAL 3	3 vials	FE0400CP1B / FE044113	2027-08-03
CAL 4	3 vials	FE0400CQ1B / FE044113	2027-08-03

Analytical Data	Specification	Result
Concentration	CAL 1 : $\leq 0.05 \mu\text{g/ml}$ CAL 2 : $0.40 - 0.60 \mu\text{g/ml}$ CAL 3 : $0.90 - 1.10 \mu\text{g/ml}$ CAL 4 : $1.35 - 1.60 \mu\text{g/ml}$	$0 \mu\text{g/ml}$ $0.49 \mu\text{g/ml}$ $0.95 \mu\text{g/ml}$ $1.39 \mu\text{g/ml}$
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	1.00 PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

06 MARS 2025

Quality Control Manager :

S. LECOURT