



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN ORGARAN CONTROL

Product Code : 223501

Lot No. : FE04621E / FE04711A

Expiration date : 2027-08-09

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE0462CL1E / FE04711A	2027-08-09
C2	6 vials	FE0462CM1E / FE04711A	2027-08-09

Analytical Data	Specification	Result
Concentration	C1 : 0.40 -0.60 U/ml C2 : 0.80 -1.20 U/ml	0.53 1.05

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval **Quality Control Approval**

Quality Control Manager :

06 MARS 2025

S. LECOURT