



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARACHIDONIC ACID
Product Code : AG003K
Lot No. : FF0468UH / FF0479
Expiration date : 2027-08-06

Components	Quantity	Lot	Expiration Date
ARACHIDONIC ACID	3 vials	FF0468UH / FF0479	2027-08-06


Analytical Data	Specification	Result
Reactivity (% AgregMax) 1.0 mM on CS or CN series	≥ 50%	90 %
Reactivity (% AgregMax) 1.0 mM on Agregometer	≥ 50%	92 %

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :



25 FEB. 2026

S.MONEUS