

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ARGATROBAN CONTROL PLASMA
Product Code : SC035K
Lot No. : FE05961C / FE062714
Expiration date : 2027-08-30

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE0596CE1C / FE062714	2027-08-30
C2	6 vials	FE0596CF1C / FE062714	2027-08-30

Analytical Data	Specification	Result
C1	0.55 - 0.80 µg/mL	0.69 µg/mL
C2	1.10 - 1.40 µg/mL	1.27 µg/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

01 AVR. 2025

S. LECOURT

