



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : GFC Control Plasmas
Product Code : SC104K
Lot No. : FE0566
Expiration date : 2027-08-24

Components	Quantity	Lot	Expiration Date
C1	2 vials	FE0566	2027-08-24
C2	2 vials	FE0566	2027-08-24
C3	2 vials	FE0566	2027-08-24

Analytical Data	Specification	Result
C1	10 - 20 min	14 min
C2	35 - 55 min	39 min
C3	65 - 110 min	96 min

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

03 AVR. 2025
S. LECOURT

Quality Control Manager :

