



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : GFC Control Plasmas
Product Code : SC104K
Lot No. : FF0727
Expiration date : 2027-08-24

Components	Quantity	Lot	Expiration Date
C1	2 vials	FE0566	2027-08-24
C2	2 vials	FE0566	2027-08-24
C3	2 vials	FF0727	2028-09-11

Analytical Data	Specification	Result
C1	10 -20 min	14 min
C2	35 - 55 min	39 min
C3	65 -110 min	68 min

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

15 AVR. 2026

S.MONEUS

