



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Item Description: LIAPHEN AT

Item Code: 120008

Lot No.: FE06063Q

Expiration date: 2027-09-07

Components	Quantity	Lot	Expiration Date
R1: Latex reagent	2 vials	FE0606IW3Q	2027-09-07
R2: Reaction Buffer	2 vials	FE0606IX3Q	2027-09-07

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	A620 (0-150%) ≥ 0.40 $R^2 \geq 0.98$ Controls measured within acceptance ranges	0.888 PASS PASS
Detection threshold	$\leq 10\%$	$\leq 7\%$
Hook effect	$>250\%$	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

Quality Control Approval

Quality Control Manager:

03 AVR. 2025

S. LECONTE