



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

Item Description : LIAPHEN Fibrinogen

Item Code : 120102

Lot No. : FE05451L / FE05701A

Expiration date : 2027-09-03

Components	Quantity	Lot	Expiration Date
R1 : Latex reagent	4 vials	FE0545JA1L / FE05701A	2027-09-03

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy	Delta A620 (0-20µg/ml) $\geq 0.40$ Controls measured within acceptance ranges $R^2 \geq 0.98$	0.65 PASS >0.99
Limit of Quantification	$\leq 1 \mu\text{g/ml}$ (in the test dilution (1:200))	PASS
Hook effect	$> 12 \text{ g/L}$ (in the test dilution (1:200))	PASS

NCI 2651 : Accepted under waiver

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval :

Quality Control Approval

Quality Control Manager :

29 AVR. 2025

S. LECOURT