



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN DABIGATRAN CONTROL PLASMA LOW
Product Code : 225001
Lot No. : FE20951A / FE212313
Expiration date : 2027-09-04

Components	Quantity	Lot	Expiration Date
C1	6 vials	FE2095DD1A / FE212313	2027-09-04
C2	6 vials	FE2095DE1A / FE212313	2027-09-04

Analytical Data	Specification	Result
Concentration	C1 : 15 -35 ng/ml C2 : 40 -80 ng/ml	27 ng/ml 63 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

15 AVR. 2025
S. LECOURT

Quality Control Manager :