


 HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN RIVAROXABAN CALIBRATOR LOW

Product Code : 226001

Lot No. : FE21101E / FE212514

Expiration date : 2027-09-06

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FE2110DJ1E / FE212514	2027-09-06
CAL II	4 vials	FE2110DK1E / FE212514	2027-09-06
CAL III	4 vials	FE2110DL1E / FE212514	2027-09-06

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -100 ng/ml	0 ng/ml 52 ng/ml 92 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

07 AVR. 2025

Quality Control Manager :

S. LECOURT

