



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : GFC Test

Product Code : CK093K

Lot No. : FE2089

Expiration date : 2027-09-03

Components	Quantity	Lot	Expiration Date
R1 : tPA with silica	3 vials	FE2089	2027-09-03
R2 : Thrombin Reagent	3 vials	FE2089	2027-09-03

Analytical Data	Specification	Result
Normal Plasma	Mean : 30 - 60 min	37.2
Controls	Lysis time within acceptance range	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

03 AVR. 2025

S. LECOURT

Quality Control Manager :