



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: Factor VIII:C Deficient Plasma
Product Code: DP040A/K
Lot No.: FE21142N / FE21271B
Expiration date: 2027-09-07

Components	Quantity	Lot	Expiration Date
Factor VIII:C Deficient Plasma	1 vial (DP040A) 6 vials (DP040K)	FE2114JV2N / FE21271B	2027-09-07

Analytical Data	Specification	Result
Calibration curve	CT (0-100%) ≥ 40 sec CT (100%) < 75 sec	73sec 57sec
Concentration FVIII	$< 1\%$	$< 1\%$
Concentration FV	$> 50\%$	77%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval:

03 AVR. 2025

S. LECOURT

Quality Control Manager: