



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** Factor XI Deficient Plasma  
**Product Code :** DP070A/K  
**Lot No. :** FE20551F  
**Expiration date :** 2027-09-07

Components	Quantity	Lot	Expiration Date
Factor XI Deficient Plasma	1 vial (DP070A) 6 vials (DP070K)	FE2055JX1F	2027-09-07

Analytical Data	Specification	Result
Calibration curve	CT (0%) >150 sec CT (100%) <100 sec	140 sec * 60 sec
Concentration FXI	< 1%	<1%
Concentration FV	> 50 %	82%

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

\* NCI 2645 : Accepted under waiver

Date of Approval :

**Quality Control Approval** Quality Control Manager :

08 AVR. 2025

S. LECOURT