



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : Factor XII Deficient Plasma
Product Code : DP080A/K
Lot No. : FE21301L / FF1024
Expiration date : 2027-09-07

Components	Quantity	Lot	Expiration Date
Factor XII Deficient Plasma	1 vial (DP080A) 6 vials (DP080K)	FE2130JY1L / FF1024	2027-09-07

Analytical Data	Specification	Result
Calibration curve	CT (0%) >180 SEC CT (100%) <65 SEC	250sec 54sec
Concentration FXII	< 1%	<1%
Concentration FV	> 50 %	85%


This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

22 AVR. 2026



S.MONEUS