

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : vWF Deficient Plasma
Product Code : DP150A/K
Lot No. : FE22191E
Expiration date : 2027-09-20

Components	Quantity	Lot	Expiration Date
vWF Deficient Plasma	1 vial (DP150A) 6 vials (DP150K)	FE22191E	2027-09-20

Analytical Data	Specification	Result
Concentration vWF	< 1 %	<1%
Concentration FVIII	< 10%	<1%
Concentration FV	> 50 %	76%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

23.06.2025

Quality Control Manager :

Claire DUNOIS
Dir, Etudes Clin. & Val.

