



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

## CERTIFICATE OF ANALYSIS

**Product Description :** ZYMUTEST HIA IgG MonoStrip

**Product Code :** RK041A

**Lot No. :** FE0512

**Expiration date :** 2027-09-17

Components	Quantity	Lot	Expiration Date
Heparin pre-coated plate	4 (x8 wells)	FE2598	2027-12-03
HIA IgG Positive Control	4 vials	FE2324	2027-12-19
Negative Control	4 vials	FE2249	2028-09-15
Anti-(h)-IgG HRP immunoconjugate	4 vials	FE2951	2028-12-21
HIA Sample diluent	2x12 ml	FE2259	2027-09-17
Wash solution	2x12 ml	FE2155	2028-09-04
Cell Lysate	4 vials	FE2323	2027-12-10
Conjugate diluent	1x10 ml	FE2156	2028-09-04
TMB	1x10 ml	240508D02	2028-05-31
Sulfuric Acid 0,45M	1x3 ml	FE2157	2028-09-05

Analytical Data	Specification	Result
Reactivity	A450 for Positive Control $\geq 1.00$ A450 for Negative Control $\leq 0.25$	PASS PASS
Pathological Plasmas	A450 $\geq 0.300$	PASS
Blank	$< 0.100$	0.018
Normal Plasma	M + 2SD $\leq 0.30$	0.14

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**      Quality Control Manager :

12 JAN. 2026  
S. LECOURT

