



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN a2 ANTIPLASMIN (LRT)
Product Code : 220502
Lot No. : FE38961K
Expiration date : 2027-10-17

Components	Quantity	Lot	Expiration Date
R1 :Plasmin	3 vials	FE3896I71K	2027-10-17
R2 : SPm-41 Substrate	3 vials	FE3896I81K	2027-10-17

Analytical Data	Specification	Result
Calibration curve : reactivity, slope, linearity and accuracy on CS	Delta A405(100-50%)> 0.05 A405/min(100%) = 0.25 ± 0.10 $R^2 \geq 0.98$ Measured values for controls in acceptance ranges	0.087 0.242 PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

12 NOV. 2025

S. LECOURT