



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CALIBRATOR
Product Code : 222301
Lot No. : FE2339PR / FE2352PS
Expiration date : 2027-10-02

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FE2339BSPR / FE2352PS	2027-10-02
CAL 2	4 vials	FE2339BTPR / FE2352PS	2027-10-02
CAL 3	4 vials	FE2339BUPR / FE2352PS	2027-10-02
CAL 4	4 vials	FE2339BVPR / FE2352PS	2027-10-02
CAL 5	4 vials	FE2339BWPR / FE2352PS	2027-10-02

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 0.05 IU/ml CAL 2 : 0.20 -0.45 IU/ml CAL 3 : 0.50 -0.80 IU/ml CAL 4 : 0.80 -1.15 IU/ml CAL 5 : 1.15 -1.50 IU/ml	0 IU/ml 0.36 IU/ml 0.72 IU/ml 1.06 IU/ml 1.41 IU/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

06 MAI 2025

S. LECOURT

