



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CONTROL LOW
Product Code : 225201
Lot No. : FE31041K
Expiration date : 2027-10-30

Components	Quantity	Lot	Expiration Date
CI	6 vials	FE3104DW1K	2027-10-30
CII	6 vials	FE3104DX1K	2027-10-30

Analytical Data	Specification	Result
Concentration	CI : 15 -35 ng/ml CII : 45 -75 ng/ml	23 ng/mL 58 ng/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval :

22 JUL. 2025

S. LECOURT

Quality Control Manager :

