



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR LOW
Product Code : 226101
Lot No. : FE24901F / FE254211
Expiration date : 2027-10-30

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FE2490DT1F / FE254211	2027-10-30
CAL II	4 vials	FE2490DU1F / FE254211	2027-10-30
CAL III	4 vials	FE2490DV1F / FE254211	2027-10-30

Analytical Data	Specification	Result
Concentration	CAL I : < 20 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 ng/mL 48 ng/mL 102 ng/mL
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **19 JUIN 2025**

Quality Control Manager :

Claire DUNOIS
Dir. Etudes Clin. & V&V

