



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ BIVALIRUDIN CALIBRATOR
Product Code : 226701
Lot No. : FE247218/FE247413
Expiration date : 2027/10/17

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FE2472cz18/FE247413	2027-10-17
CAL 2	2 vials	FE2472da18/FE247413	2027-10-17
CAL 3	2 vials	FE2472db18/FE247413	2027-10-17
CAL 4	2 vials	FE2472dc18/FE247413	2027-10-17
CAL 5	2 vials	FE2472dd18/FE247413	2027-10-17

Analytical Data	Specification	Result
Concentration	CAL 1 : < 0.25 µg/mL CAL 2 : 0.8 -1.2 µg/mL CAL 3 : 1.6 -2.4 µg/mL CAL 4 : 3.0 - 4.0 µg/mL CAL 5: 4.0 - 5.0 µg/mL	0.00 µg/mL 1.10 µg/mL 2.22 µg/mL 3.57 µg/mL 4.77 µg/mL
Accuracy of calibration with values assigned	Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval

Date of Approval : 27 MAI 2025
S. LECOURT

Quality Control Manager :

