



A Company of Sysmex Group



HYPHEN BioMed  
155 rue d'Eragny  
95000 Neuville-sur-Oise  
France

### CERTIFICATE OF ANALYSIS

**Product Description :** CEPHEN 2.5 LS  
**Product Code :** CK522K  
**Lot No. :** FE3719SK / FE3733  
**Expiration date :** 2027-10-10

Components	Quantity	Lot	Expiration Date
CEPHEN 2.5 LS	6 vials	FE3719KDSK / FE3733	2027-10-10

Analytical Data	Specification	Result
aPTT on Normal Plasma	Ratio < 1.20	1.08 on STAR 1.04 on CS 1.07 on CN

*This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.*

Date of Approval : **Quality Control Approval**

**12 NOV. 2025**  
**S. LECOURT**

Quality Control Manager :