



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST Fibrinogen

Product Code : RK024A

Lot No. : FE2448

Expiration date : 2027-10-22

Components	Quantity	Lot	Expiration Date
pre-coated plate	12x8 wells	FE2513	2027-10-29
Fibrinogen Standard	3 vials	FE2446	2027-10-23
Fibrinogen Control I	1 vial	FE2445	2027-10-23
Fibrinogen Control II	1 vial	FE2447	2027-10-23
Fibrinogen HRP immunoconjugate	3 vials	FE2684	2027-11-19
B2F Sample diluent	2x50 ml	FE2451	2027-10-22
Wash solution	1x50 ml	FE0010	2028-06-15
B2F Conjugate diluent	1x25 ml	FE2452	2027-10-22
TMB	1x25 ml	250104D01	2029-01-31
Sulfuric Acid 0,45M	1x6 ml	FE0018	2028-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 40 ng/ml $\geq$ 1.45	1.67
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.150	0.014
Detection threshold	$\leq$ 1 ng/ml	< 1 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : 17 JUIN 2025

Quality Control Manager : Isabelle CORNUEJOLS

p.o Resp. Labo CQ

