



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description: BIOPHEN LMWH CONTROLC3
Product Code: 223801
Lot No.: FE2630QG / FE2652
Expiration date: 2027-11-14

Components	Quantity	Lot	Expiration Date
C3	12 vials	FE2630BOQG / FE2652	2027-11-14

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml	0.74 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

19 JUN 2025

Quality Control Manager:

Claire DUNOIS
Dir. Etudes Clin. & Val.