



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN UFH CONTROL C2
Product Code : 223901
Lot No. : FE2645QJ / FE2655
Expiration date : 2027-11-14

Components	Quantity	Lot	Expiration Date
C2	12 vials	FE2645BYQJ / FE2655	2027-11-14

Analytical Data	Specification	Result
Concentration	C2 : 0.40 -0.60 IU/mL	0.50 IU/mL

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

18.06.2025

Quality Control Manager :

Claire DUNOIS
Dir. Etudes Clin. & Val.