



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL LOW CI
Product Code : 224301
Lot No. : FE2576QC / FE2611QH
Expiration date : 2027-11-08

Components	Quantity	Lot	Expiration Date
CI	12 vials	FE2576BQQC / FE2611QH	2027-11-08

Analytical Data	Specification	Result
Concentration	CI : 0.15 -0.35 IU/ml	0.29 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

19 JUN 2025

Quality Control Manager :

Claire DUNOIS
Dir. Etudes Clin. & Val.