



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN APIXABAN CALIBRATOR

Product Code : 226201

Lot No. : FE25391G / FE254611

Expiration date : 2027-11-02

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FE2539DO1G / FE254611	2027-11-02
CAL 2	4 vials	FE2539DP1G / FE254611	2027-11-02
CAL 3	4 vials	FE2539DQ1G / FE254611	2027-11-02

Analytical Data	Specification	Result
Concentration	CAL 1 : ≤ 50 ng/ml CAL 2 : 200 -400 ng/ml CAL 3 : 400 -600 ng/ml	0 ng/ml 317 ng/ml 516 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **19 JUIN 2025**

Quality Control Manager : **Claire DUNOIS**
Dir. Etudes Clin. & Val.

