



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : HEMOCLOT THROMBIN INHIBITORS

Product Code : CK002L

Lot No. : FE2741QP / FE2785

Expiration date : 2027-11-29

Components	Quantity	Lot	Expiration Date
R1: Normal Pooled Plasma	3 vials	FE2741APQP / FE2785	2027-11-29
R2 : Human Calcium Thrombin	3 vials	FE2741AQQP / FE2785	2027-12-12

Analytical Data	Specification	Result
Dabigatran standard calibration curve on STAR	Delta CT (50-500 ng / mL) : 50 ± 10 sec	50.7 sec
	$R^2 \geq 0.98$	PASS
	Controls measured in acceptance ranges	PASS
Normal Plasma	Dabigatran conc. < 10 ng / mL	PASS

Calibration Curve for information

Standard Range

Dabigatran Conc.-ng/mL	0	270	544
CT (Sec.) STAR	28.5	58.6	89.7
CT (Sec.) CS	25.9	77.1	120.3

QC:	TV* (ng/ml)	MV* (ng/ml)
C1	[82-122]	100
C2	[213-289]	244

Low Range

Dabigatran Conc.-ng/mL	0	43	110
CT (Sec.) STAR	28.0	48.5	79.1
CT (Sec.) CS	41.1	57.8	89.2

QC:	TV* (ng/ml)	MV* (ng/ml)
CI	[17-37]	24
CII	[50-76]	59

* TV= Target Value - MV= Measured Value

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Manager :

Quality Control Approval

09 JUL. 2025

S. LECOURT